

Please read this package insert carefully before use.

Capilia™ Flu Neo

INTENDED USE

To detect influenza A virus antigens and influenza B virus antigens in nasal aspirate, nasal swab, nasal discharge/nasal mucus or pharyngeal swab (to assist in the diagnosis of influenza virus infectious disease).

SUMMARY AND EXPLANATION OF THE TEST

Influenza spreads around the world in seasonal epidemics, resulting in about 3 to 5 million annual cases of severe illness and about 290,000 to 650,000 annual deaths, rising to millions in some pandemic years. (WHO News Release, 2017)

Common symptoms are chills, fever, sore throat, muscle pains, headache, coughing, weakness/fatigue and general discomfort.

Diagnosis is difficult because the initial symptoms can be similar to those caused by other infectious agents.

The influenza A virus is usually more prevalent and is associated with the most serious influenza epidemics, while influenza B infections usually present more mild symptoms.

Because the influenza virus is highly contagious, rapid diagnosis and prompt treatment can have a positive effect on public health. And the ability to distinguish between A or B antigens can help the physician to prescribe an appropriate antiviral therapy.


Administration of antiviral therapy within 48 hours of symptom onset is recommended for more rapid reduction of symptoms and to reduce viral shedding.

Capilia Flu Neo can provide rapid and accurate detection of influenza A and/or B virus antigens from symptomatic patients. No special instruments or equipment are required.

PRINCIPLE OF THE TEST

Measurement using this product is based on an immunochromatography assay using a monoclonal antibody that recognizes influenza virus antigens.

This product comprises a test plate with a carrier strip containing a sample placement area, a reagent area including a colloidal platinum-gold labeled anti-influenza A and B virus monoclonal antibody (mouse) (hereinafter referred to as "colloidal platinum-gold labeled antibody"), a reading area [A] that fixes the anti-influenza A virus monoclonal antibody (mouse) (hereinafter referred to as "anti-influenza A virus antibody"), a reading area [B] that fixes the anti-influenza B virus monoclonal antibody (mouse) (hereinafter referred to as "anti-influenza B virus antibody"), and a reading area [C] that fixes the anti-mouse immunoglobulin polyclonal antibody (rabbit) (hereinafter referred to as "anti-mouse immunoglobulin antibody").



When a sample is placed on the sample placement area of the test plate, the colloidal platinum-gold labeled antibody dissolves and forms an immune complex with the influenza A and/or B virus antigens in the sample. This immune complex migrates through the developing area by capillary action, is captured by the anti-influenza A virus antibody and/or the anti-influenza B virus antibody fixed in the developing area, and forms a black line of colloidal platinum-gold in the reading area [A] and/or [B]. The black line visually displays the existence of influenza virus antigens in the sample.

Regardless of the existence of influenza virus antigens in the sample, excess colloidal platinum-gold labeled antibodies further migrate through the developing area, are captured by anti-mouse immunoglobulin antibodies fixed in the developing area, and form a black line in the reading area [C]. This means the colloidal platinum-gold labeled antibodies have migrated normally.

REAGENTS AND MATERIALS PROVIDED

REF CAFL0570 Capilia Flu Neo (20 Tests)

Test plates

• Components

Colloidal platinum-gold labeled anti-influenza A, B virus monoclonal antibody (mouse)

Anti-influenza A virus monoclonal antibody (mouse)

Anti-influenza B virus monoclonal antibody (mouse)

Extraction Buffer (to be used equally with the four products.)^{Note}

• Components

Buffer, detergent, sodium azide (0.09%)

Nozzles

REF CAFL0571 Capilia Flu Neo (10 Tests)

Test plates

• Components

Colloidal platinum-gold labeled anti-influenza A, B virus monoclonal antibody (mouse)

Anti-influenza A virus monoclonal antibody (mouse)

Anti-influenza B virus monoclonal antibody (mouse)

Extraction Buffer (to be used equally with the four products.)^{Note}

• Components

Buffer, detergent, sodium azide (0.09%)

^{Note} The extraction buffer is able to be used equally with the four products below:

- Capilia Flu Neo (rapid test for detecting influenza virus antigen)
- Capilia Adeno Neo (rapid test for detecting adenovirus antigen)
- Capilia RSV Neo (rapid test for detecting RS virus antigen)
- Capilia hMPV (rapid test for detecting human metapneumovirus antigen)

Specimen	Flu	Adeno	RSV	hMPV
Nasal swab	○	○	○	○
Nasal aspirate	○	○	○	○
Nasal discharge/Nasal mucus	○	×	×	×
Pharyngeal swab	○	○	×	○
Keratoconjunctivitis swab	×	○	×	×

Nozzles

Nasal Swabs <See Item 1) of "3. Precautions when using swabs.">

MATERIALS REQUIRED BUT NOT PROVIDED

REF CAFL0570 Capilia Flu Neo (20 Tests)

Timer, micropipette, pipette tips, suction machine, suction trap, the specimen collecting sheet for nasal discharge, sterile swabs (as listed here)

Recommended Swab

The following swabs are recommended for use with the kit.

For Nasal Swab

- ① FLOQSwabs™ (Cat No. 534CS01, Copan Italia S.p.A, Italy)
- ② Sterilized Swab P156A 10 pcs (Cat No. 4124, HEIWA MEDIC. CO., LTD, Japan)

For Pharyngeal Swab

- ① FLOQSwabs™ (Cat No. 502CS01, Copan Italia S.p.A, Italy)
- ② Sterilized Swab PL6S 10 pcs (Cat No. 4371, HEIWA MEDIC. CO., LTD, Japan)

To collect nasal aspirate, any of the above swabs may be used.

Acceptable Swab

The following swabs are acceptable for use with the kit.

- Tip material
Rayon, flocked nylon and polyester
- Standard tip size
For nasal swab
Plain dry swab: maximum diameter 3 mm, length 12 mm
Flocked swab: maximum diameter 3 mm, length 15 mm

For pharyngeal swab

- Plain dry swab: maximum diameter 6 mm, length 14 mm
- Flocked swab: maximum diameter 6 mm, length 16 mm
- Shaft material
Paper, plastic (PS, nylon), aluminum

Unacceptable Swab

Do not use calcium alginate swabs.

REF CAFL0571 Capilia Flu Neo (10 Tests)

Timer, micropipette, pipette tips, suction machine, suction trap, the specimen collecting sheet for nasal discharge

WARNING AND PRECAUTIONS

1. Precautions when handling (including hazard control)


- 1) All specimens should be handled as potentially infectious, and special precautions should be exercised.
- 2) In consideration of the risk of infection, wear protective clothes such as a mask and gloves and handle the specimens and samples carefully during the test.
- 3) If the extraction buffer gets into your eyes, immediately flush with a large quantity of water for 15 minutes or more. If you still feel some abnormality, see a doctor for treatment.
- 4) If the extraction buffer comes into contact with your hands or clothes, wash your hands and/or clothes with soap and a large quantity of water.

2. Precautions when using

- 1) This product reacts only with influenza A and B viruses and does not react with C virus.
- 2) This product is a rapid test for detecting influenza A and B virus antigen. **A definite diagnosis should be made by an attending physician, in combination with the clinical symptoms, the result of viral isolation culture test and other test results.**
- 3) This product should be used in accordance with the procedure stated in the package insert.
- 4) In order to prevent deterioration, this product should be stored between 2°C and 30°C, avoiding high temperatures, high humidity and direct sunlight.
- 5) If this product has been refrigerated, it must be removed from the refrigerator at least 30 minutes before use and kept at room temperature when used for testing.
- 6) **The aluminum pouch containing a test plate should not be opened until the test plate is about to be used.**
- 7) The sample placement area and the reading area of the test plate should not be touched with the hands.
- 8) A precipitate may be seen in the extraction buffer, but the product can be used as it is, because the precipitate has been shown not to affect test results.
- 9) For nasal sampling, do not keep forcibly insert the swab, when the distance to the site is clearly shorter than usual. In particular, there is the possibility of resistance being imposed on the stick when the sample is collected from an infant or a patient with a narrow nasal cavity. In such a case, do not swab hard, exerting force on the stick. Moreover, do not rotate the stick forcibly.
- 10) For nasal sampling, any mass of mucus on the tip of the swab should be gently removed with gauze. Do not wipe the tip too hard. Mucosal epidermal cells should remain on the tip for testing.
- 11) When sampling a nasal discharge, be careful not to spill or scatter the samples, as it may cause a secondary infection.
- 12) Do not use any products beyond the expiration date.

3. Precautions when using swabs

- 1) Use the swab provided in the kit according to the following instructions:
Open the peel package from the side indicating "◀ 開封口 (PEEL HERE)" and pull the swab out.
Peel the film, then remove the swab by pinching the shaft, avoiding contact with the swab tip, and directly use the swab for collecting the specimen.



- 2) A swab is for single use only; reuse is not allowed in any case. Reuse may cause a risk of infection and/or inaccurate results.

- 3) Do not re-sterilize unused swabs.
- 4) Do not re-pack.
- 5) To be handled by trained personnel only.
- 6) Swabs should not be used if the swab package is open or damaged and collect samples immediately after opening.
- 7) Do not use a swab if it is broken, bent or stained.
- 8) Avoid the following when using the swab since they may cause the breakage of the shaft.
 - To use the swab in a manner such that the shaft (especially the narrow part of the shaft) receives too much force, pressure or torsional load at the time of specimen collection.
 - To intentionally deform the shaft, more specifically, to bend, arch or break the shaft.
- 9) Particularly when collecting a pharyngeal swab from a child, collect the specimen with great caution lest he or she should act up, bite or break the swab and swallow the broken part, and by firmly retaining the upper body of the child to be tested

4. Precautions for disposal

- 1) Because used test plates, swabs, tubes and nozzles after use, remaining samples, etc. may cause infections, they should be autoclaved (121°C, 20 min) or soaked in 0.1% sodium hypochlorite for more than one hour. When remaining reagents or their accessories are disposed of, they should be treated in accordance with the laws and regulations concerning medical waste disposal and water pollution control.
- 2) In the extraction buffer, 0.09% of sodium azide is included as a preservative. When solutions containing sodium azide continue to be discarded over a long period of time, explosive metallic azide may be produced if a drain is made of metal. Therefore, they should be discarded with a large quantity of water.

STORAGE CONDITIONS


Storage : Store at 2°C to 30°C. **DO NOT FREEZE.**

Keep away from direct sunlight.

Do not use test plate or extraction buffer after expiration date.


SPECIMEN COLLECTION AND PREPARATION

1. Methods of specimen collection




1) Sampling of nasal aspirate

Firmly insert one tube of the suction trap into the suction pump, and the other tube into a nasal cavity through an external nostril. Collect the nasal discharge aspirate in the suction trap by operating the suction pump. Soak a swab in the nasal aspirate collected by the trap, and let the swab absorb the nasal aspirate well. When nasal aspirate is taken using a micropipette or other instruments, dilute the nasal aspirate twofold with physiological saline and sample 200 µL of this dilution.



2) Sampling of nasal swab

Firmly insert a nasal swab into the nasal cavity and collect mucosal epithelium by swabbing the nasal turbinate several times.



3) Sampling of pharyngeal swab

Firmly insert a pharyngeal swab into the pharynx through the oral cavity, and collect the mucosal epithelium by swabbing the posterior wall of the pharynx and the palatine tonsil several times, centering around the rubefacient portion. **Avoid touching saliva.**

If the specimen is mixed with saliva, the lines on the test plate may become fainter.




4) Sampling of nasal discharge/nasal mucus

When the patient's condition is judged to be suitable for collecting nasal discharge, the patient should blow his/her nose using the specimen collecting sheet. Swab the collected nasal discharge with a nasal swab. Or directly swab the nostril to obtain the specimen.


If either sampling method collects insufficient samples, try other methods.

With a collected specimen, prepare the sample and perform the test as quickly as possible.

2. Sample preparation



Remove the aluminum sealing cap from the extraction buffer tube, while taking care not to spill the liquid.



Soak the swab that collected the specimen in the extraction buffer, and stir well.


Then, pinch the tip of the swab firmly with the soft wall of the extraction buffer tube with your fingers and squeeze out the swab. Use this squeezed-out liquid as the sample.

When using a nasal aspirate specimen diluted twofold with physiological saline, add 200 μL of the specimen to the extraction buffer in the tube, and mix well. Use this mixture as the sample.

3. Precautions for sample preparation

For a highly viscous sample that can cause filter clogging, dilute the sample twofold with physiological saline before use.

TEST PROCEDURE



- 1) Firmly attach the nozzle (with a filter) provided in the kit to the top of the extraction buffer tube.
- 2) Hold the middle of the tube with the fingers and dispense **3 drops of the sample (80–120 μL)** onto the sample placement area of the test plate. Hold the tube perpendicularly and take care not to let the tip of the nozzle touch the sample placement area.
- 3) Observe the reading area of the test plate after **3 to 5 minutes for a positive result and after 5 minutes for a negative result**, and interpret the result according to the "READING TEST RESULTS."

Precautions for sample placement

If an excessive amount of the sample is dispensed, the reaction time may be prolonged due to the dilution of the colloidal platinum-gold labeled antibody based on the principle of the test, and no line may be seen at [C] and/or [T] in the reading area within the judgment time or the line is faint.

READING TEST RESULTS

Allow the samples to react according to the test procedure and read the black lines that appear in the reading area.

A-positive



When black lines are seen at both [A] and [C] in the reading area (two lines), the result is read as positive for influenza A virus antigen. When a very faint black line is seen in the reading area [A], the result is interpreted as positive.

B-positive



When black lines are seen at both [B] and [C] in the reading area (two lines), the result is read as positive for influenza B virus antigen. When a very faint black line is seen in the reading area [B], the result is interpreted as positive.

Negative



When no black line is seen at [A] and [B] in the reading area but a black line is seen only at [C] in the reading area (one line), the result is read as negative. When the line at [C] in the reading area is faint but visually recognizable, chromatographic development has occurred normally.

Retesting



When no black line is seen at [C] in the reading area, there may be some problem with the test procedure or the reagent quality. The test should be performed again, using another test plate.



A line that appears anywhere within the sections of the reading area, which are separated by color, is considered valid.

(Note)

1. When black lines are seen at both at [A] or [B] and [C] in the reading area 3 to 5 minutes after dispensing the sample, the result is read as A-positive or B-positive. No black line at [A] and [B] in the reading area even 5 minutes after dispensing the sample indicates a negative result.
2. Do not use the test plate for reading a result beyond the judgment time as the result may change due to drying, etc.
If the amount of antigen in the sample is very low and close to the detection limit of this product, a black line may appear at [A] or [B] in the reading area after 5 minutes or more of the judgment time, due to the feature of immunochromatography. In addition, due to a non-specific reaction caused by specimen-derived components, a black line may appear at [A] and/or [B] in the reading area after 5 minutes or more of the judgment time in rare cases. A definite diagnosis should be made comprehensively, not only based on the test result of this product but also taking other test results and clinical symptoms into account.
3. An A-positive result does not rule out the presence of B-infection. Contrarily, a B-positive result does not rule out the presence of A-infection. On rare occasions, the result shows positivity for both A and B.
4. If the amount of antigen is very high, a very thick line may be seen at [A] or [B] in the reading and no black line may be seen at [C] in the reading area. In that case, dilute the sample with more extraction buffer and perform the test again.
Example) Method for dilution of sample : Dispense 3 drops of the sample to a new extraction buffer tube, mix thoroughly and use the solution as the test sample.
5. The line is valid even if there is unevenness in depth and there are breaks in the line.

LIMITATIONS

1. This product is a rapid test for detecting influenza A and B virus antigen.
A definite diagnosis should be made by an attending physician, in combination with the clinical symptoms, the result of viral isolation culture test and other test results.
2. If a pharyngeal swab, nasal discharge or nasal mucus is used as a specimen, pay special attention to the method of collection, as the test tends to be less sensitive than those of nasal swabs and aspirates.
3. The test plate should be used immediately after opening the packaging.
When it absorbs moisture, the quality deteriorates and an accurate result cannot be obtained.
4. This product should be used for *in vitro* diagnosis only and should not be used for any other purposes.
5. Please use this product following the operational method described in this package insert. We cannot guarantee results obtained from any other operations and for any other purposes that are not described in the package insert.
6. The extraction buffer contains sodium azide. If the solution comes into contact with the eye or mouth or adheres to the skin by mistake, take emergency measures such as thorough washing with water and seek medical treatment, if necessary.

PERFORMANCE CHARACTERISTICS

1. Clinical data

The result of the clinical performance evaluation in Japan (Comparison with isolation culture method).

Nasal swab, nasal aspirate, pharyngeal swab : Exams carried out in the flu season in 2006 to 2007.

Nasal discharge/nasal mucus : Exams carried out in 2008.

Kind of samples	Type	Sensitivity (%)	Specificity (%)	Accuracy (%)	Total number
Nasal swab	Type A	94.3 (100/106)	95.4 (187/196)	95 (287/302)	302
	Type B	100 (69/69)	98.7 (230/233)	99 (299/302)	302
Nasal aspirate	Type A	95.1 (58/61)	98.4 (188/191)	97.6 (246/252)	252
	Type B	100 (61/61)	100 (191/191)	100 (252/252)	252
Pharyngeal swab	Type A	87.3 (55/63)	94.8 (128/135)	92.4 (183/198)	198
	Type B	91.5 (75/82)	99.1 (115/116)	96 (190/198)	198
Nasal discharge/nasal mucus	Type A	84.1 (53/63)	96.9 (94/97)	91.9 (147/160)	160
	Type B	96.2 (50/52)	98.1 (106/108)	97.5 (156/160)	160

2. Sensitivity (Detection limit)

The minimum detection limit is 7.5×10^3 TCID₅₀/test for the influenza A virus antigen and is 7.5×10^4 TCID₅₀/test for the influenza B virus antigen.

3. Reactivity

Reactivity was found in the following strains :

1) Human origin type A virus

A/New Jersey/8/76 (H1N1)
 A/Sendai/782/06 (H1N1) A/Sendai/197/07 (H1N1) A/Adachi/1/57 (H2N2)
 A/Sendai/F492/06 (H3N2) A/Sendai/958/07 (H3N2) A/Aichi/2/68 (H3N2)
 A/Anhui/1/2013 (H7N9)

2) Human origin A(H1N1)pdm09

A/Osaka/50/09 A/Osaka/51/09 A/Osaka/52/09 A/Osaka/55/09
 A/Osaka/56/09 A/Osaka/57/09 A/Osaka/58/09 A/Osaka/59/09
 A/Osaka/60/09 A/Osaka/61/09 A/Osaka/63/09 A/Osaka/64/09
 A/Osaka/65/09 A/Osaka/66/09 A/Osaka/69/09 A/Osaka/70/09
 A/Osaka/71/09 A/Osaka/72/09 A/Osaka/78/09 A/Osaka/83/09
 A/Osaka/84/09 A/Osaka/85/09 A/Osaka/90/09 A/Osaka/91/09
 A/Osaka/100/09 A/Osaka/101/09 A/Osaka/102/09 A/Osaka/103/09
 A/Osaka/104/09 A/Osaka/105/09 A/Osaka/106/09 A/Osaka/107/09
 A/Osaka/108/09 A/Osaka/109/09 A/Osaka/110/09 A/Osaka/112/09
 A/Osaka/114/09 A/Osaka/115/09 A/Osaka/116/09 A/Osaka/118/09
 A/Osaka/119/09 A/Osaka/126/09 A/Osaka/130/09 A/Osaka/139/09
 A/Osaka/143/09 A/Osaka/144/09 A/Osaka/146/09 A/Osaka/148/09
 A/Osaka/157/09 A/Osaka/164/09 A/Osaka/165/09 A/Osaka/167/09
 A/Osaka/168/09 A/Osaka/169/09 A/Osaka/171/09 A/Osaka/172/09
 A/Osaka/174/09 A/Osaka/176/09 A/Osaka/193/09

3) Type A virus of other than human origin

A/duck/Tottori/723/80 (H1N1) A/duck/Hokkaido/17/01 (H2N3)
 A/duck/Mongolia/4/03 (H3N8) A/duck/Czechoslovakia/1/56 (H4N6)
 A/chicken/Yamaguchi/7/04 (H5N1) A/whooper swan/Hokkaido/1/08 (H5N1)
 A/whooper swan/Mongolia/3/05 (H5N1) A/duck/Pennsylvania/10218/84 (H5N2)
 A/duck/HongKong/820/80 (H5N3) A/turkey/Massachusetts/3740/65 (H6N2)
 A/shearwater/Australia/1/72 (H6N5) A/chicken/Italy/99 (H7N1)
 A/chicken/Pakistan/447/95 (H7N3) A/seal/Massachusetts/1/80 (H7N7)
 A/chicken/Netherlands/2586/03 (H7N7) A/tufted duck/Shimane/124R/80 (H7N7)
 A/duck/Mongolia/119/2008 (H7N9) A/duck/Mongolia/129/2010 (H7N9)
 A/turkey/Ontario/67 (H8N4) A/turkey/Ontario/6118/68 (H8N4)
 A/turkey/Wisconsin/66 (H9N2) A/chicken/Germany/N/49 (H10N7)
 A/duck/England/1/56 (H11N6) A/duck/Alberta/60/76 (H12N5)
 A/gull/Maryland/704/77 (H13N6) A/mallard/Astrakhan/263/82 (H14N5)
 A/duck/Australia/341/83 (H15N8) A/black-headed gull/Sweden/5/99 (H16N3)
 A/swine/Iowa/15/30 (H1N1) A/swine/Niigata/1/77 (H1N1)
 A/swine/Niigata/1/78 (H1N1) A/swine/Toyama/1/78 (H1N1)
 A/swine/Kanagawa/1/78 (H1N1) A/swine/Shizuoka/1/78 (H1N1)
 A/swine/Shimane/1/78 (H1N1) A/swine/Hokkaido/80 (H1N1)
 A/swine/Hokkaido/2/81 (H1N1) A/swine/Saitama/96 (H1N2)
 A/swine/Miyagi/5/03 (H1N2) A/swine/Hong Kong/126/82 (H3N2)
 A/swine/Obihiro/10/85 (H3N2) A/swine/Chonburi/02 (H3N2)

4) Human origin type B virus

B/Sendai/1708/05 B/Sendai/942/07 B/Lee/40

4. Cross reactivity

No cross-reactivity was found in all the viruses and bacteria listed below.

1) Viruses

Adenovirus Type 1-6, 11	Influenza virus C
Parainfluenza virus Type 1-4	Respiratory syncytial virus (A) (B)
Rhinovirus Type 2	Coxsackie virus Type A9, A16, B1-6
Echovirus Type 4, 6, 9, 11, 14, 16	Cytomegalovirus
Human Metapneumovirus	

2) Bacteria

<i>Acinetobacter baumannii</i>	<i>Bacillus cereus</i>
<i>Bacteroides fragilis</i>	<i>Bordetella pertussis</i>
<i>Branhamella catarrhalis</i>	<i>Campylobacter jejuni</i>
<i>Citrobacter freundii</i>	<i>Eikenella corrodens</i>
<i>Enterococcus faecalis</i>	<i>Enterobacter cloacae</i>
<i>Fusobacterium nucleatum</i>	<i>Gardnerella vaginalis</i>
<i>Haemophilus influenzae</i>	<i>Haemophilus parainfluenzae</i>
<i>Kingella kingae</i>	<i>Klebsiella oxytoca</i>
<i>Lactobacillus casei</i>	<i>Mycobacterium abscessus</i>
<i>Mycobacterium avium</i>	<i>Mycobacterium intracellulare</i>
<i>Mycobacterium tuberculosis</i>	<i>Neisseria meningitidis</i>
<i>Nocardia asteroides</i>	<i>Pasteurella multocida</i>
<i>Peptostreptococcus anaerobius</i>	<i>Porphyromonas asaccharolyticus</i>
<i>Prevotella intermedia</i>	<i>Prevotella melaninogenica</i>
<i>Salmonella choleraesuis</i> (sub, minnesota)	<i>Serratia marcescens</i>
<i>Staphylococcus aureus</i>	<i>Staphylococcus epidermidis</i>
<i>Streptococcus bovis</i> (II Group D)	<i>Streptococcus</i> sp. group A, B, C, F, G
<i>Streptococcus milleri</i>	<i>Streptococcus mutans</i>

Streptococcus oralis
Streptococcus sanguis
Chlamydomytila psittaci

Streptococcus pneumoniae
Chlamydomytila pneumoniae

INTERFERING SUBSTANCES

The following substances were found to have no effect on the results at the concentrations indicated.

Whole blood (0.25%), acetylsalicylic acid (20 mg/mL), ambroxol hydrochloride (375 ng/mL), dequalinium hydrochloride (6.25 ng/mL), oxymetazoline hydrochloride (100 ng/mL), dried Platycodon extract (555 ng/mL), disodium cromoglycate (5 mg/mL), zanamivir (500 ng/mL), diphenhydramine hydrochloride (10 mg/mL), cyproheptadine hydrochloride hydrate (200 ng/mL), cefixime (2.5 mg/mL), dextromethorphan hydrobromide monohydrate (10 mg/mL), Naphazoline nitrate (125 ng/mL), (R) - (-) -phenylephrine hydrochloride (1 mg/mL), fluticasone propionate (127.5 ng/mL), chlorpheniramine maleate (5 mg/mL)

REFERENCES

- 1) Kudo K, Takasaki J, Manabe T, Uryu H, Yamada R, et al. Systemic Corticosteroids and Early Administration of Antiviral Agents for Pneumonia with Acute Wheezing due to Influenza A (H1N1) pdm09 in Japan. *PLoS ONE*. 2012;7(2):e32280.
- 2) Iwaki N, et al. Review on the Epidemic of Influenza from 2009 to 2010 in Japan. *Jpn J Clin Exp Med*. 2010;87:1489-1499.
- 3) Kurita I, Kamiya C, Kouga K, Amano T. Evaluation of rapid influenza test kit and detection of influenza A (H1N1) pdm09 using RT-PCR. *J Clin Lab Inst Reag*. 2010;33(5):645-648.
- 4) Takasaki Y, et al. Evaluation of Rapid Influenza Test Kit "ImunoAce Flu" Using Pt-Au Colloid. *Jpn J Clin Exp Med*. 2008;85:1804-1807.
















INQUIRES

 **TAUNS Laboratories, Inc.**
 761-1, Kamishima, Izunokuni,
 Shizuoka, 410-2325 Japan

FAX : +81-558-76-0022

 **Emergo Europe**
 Prinsessegracht 20
 2514 AP The Hague
 The Netherlands

GLOSSARY OF SYMBOLS

- | | | | |
|---|------------------------------------|---|---|
|  | CE mark of conformity |  | Authorized representative in the European Community |
|  | In vitro diagnostic medical device |  | Do not reuse |
|  | Temperature limitation |  | Manufacturer/Manufactured by |
|  | Use by YYYY-MM |  | Consult instructions for use |
|  | Batch code |  | Caution, consult accompanying documents |
|  | Catalog number |  | Keep away from sunlight |
|  | Contents sufficient for <n> tests |  | Fragile, handle with care |
|  | Open here | | |