




# HPV 14 Screening & 16,18, 45 Typing Real-TM Quant HANDBOOK

Real Time PCR kit for quantitative detection and genotyping of *Human Papillomavirus* types 16,18,45 and quantitative detection of HPV types 31, 33, 35, 39, 51, 52, 56, 58, 59, 66, 68

**REF** V31-100/F FRT

 100

## NAME

### HPV 14 Screening & 16,18,45 Typing Real-TM Quant

## INTRODUCTION

Genital infection with HPV is one of the most common sexually transmitted diseases (STDs) of viral etiology worldwide (20% - 46% in different countries in sexually active young women).

Cervical cancer is the second most common cancer in women worldwide, and a compelling body of clinical, epidemiological, molecular, and experimental evidence has established the etiological relationship between some sexually transmitted HPV genotypes and cervical neoplasia throughout the world. Based on the frequency of detection of HPV genotypes from different grades of Cervical Intraepithelial Neoplasia (CIN Grades I – III), HPV genotypes are subdivided into High-risk HPV types (16, 18, 31 and 45), Intermediate-risk types (33, 35, 39, 51, 52, 56, 58, 59, and 68), and Low-risk types (6, 11, 42-44).

Several methods have been used to diagnose clinical or subclinical infection with HPVs including clinical observation, cytological screening by Pap smear, electron microscopy, immunocytochemistry, but these methods have some disadvantages such as non-standardization and subjectivity, insufficient sensitivity and low predictable value. The most perspective way of HPV diagnosis is a direct detection of DNA of the human papilloma virus of high carcinogenic risk by the polymerase chain reaction. While the value of the Pap smear in routine screening for cervical dysplasia is undisputed, it is now known that 99% of cases of cervical carcinoma are caused by infection with twelve genotypes of the human papilloma virus (HPV). Identification of these high-risk genotypes is very valuable in the management of cervical carcinoma, both as a prognostic indicator and as a secondary screening test where results of a Pap smear are inconclusive. Results from the combination of the Pap smear and the HPV DNA test can aid in determining the intervals for screening.

The PCR-based methods have been used successfully for the detection and typing of genital HPV genotypes in clinical specimens such as cervical swabs or scrapes, cervicovaginal lavages, frozen biopsies and formalin-fixed paraffin-embedded tissues.

## INTENDED USE

Kit **HPV 14 Screening & 16,18,45 Typing Real-TM Quant** is an *in vitro* Real Time amplification test for quantitative genotyping of *Human Papillomavirus DNA* types 16,18,45 and simultaneous quantitative detection of HPV DNA 31,33,35,39,51,52,56,58,59,66,68 (total 14 HPV genotypes detected) in the urogenital swabs. It being known, that the parameter of viral load has a prognostic value and the viral load less than  $10^5$  HPV genomic equivalents in the swab or  $10^3$  genomic equivalents for  $10^5$  cells is considered as insignificant and indicates the presence of transitory infection, however such level of load may have a value only in cases of treatment monitoring. Viral load of more than  $10^5$  genomic equivalents for  $10^5$  cells is considered to be important with high significance and indicates the existence of dysplastic changes or high risk of their occurrence. Quantitative detection of viral load allows to evaluate the character of the infection and to make a forecast concerning the stage of the disease.

HPV 14 Screening & 16,18,45 Typing Quant detect the most widespread and oncogenic 14 genotypes of human papilloma virus with determination of clinical significance. Since the human papilloma virus is an intracellular agent, there is need to monitor the presence of cellular material in the sample, in order to avoid false-negative results. HPV 14 Screening & 16,18,45 Typing Quant kit contains the internal control (human beta-globine gene), which allows to control the presence of cellular material in the sample.

## PRINCIPLE OF ASSAY

Kit **HPV 14 Screening & 16,18,45 Typing Real-TM Quant** is based on two major processes: isolation of DNA from specimens and Real Time amplification. It is an in vitro Real Time amplification test for quantitative detection and genotyping of HPV 16,18,45 and simultaneous quantitative detection of HPV types 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59, 66, 68 (total 14 genotypes detected).

If the swab is not correctly prepared (high quantity of mucous or insufficient quantity of epithelial cells) the Internal Control will not be detected. The kit contains the quantitative standards with known concentration of HPV DNA which allows to determinate the viral load. For the calculation of viral load it is used the relation between the obtained HPV DNA concentration and the quantity of genomic DNA which allows to eliminate the possible errors during the sample preparation.

## MATERIALS PROVIDED

### Module No.1: Real Time PCR kit (V31-100/F FRT)

Part N° 2 – “HPV 14 Screening & 16,18,45 Typing Real-TM Quant”:

- **PCR-mix-1-FRT HPV14**, 1 x 1,2 ml;
- **PCR-mix-2 buffer**, 1 x 0,60 ml
- **TaqF DNA Polymerase**, 1 x 0,06 ml
- **Negative Control**, 1,2 ml\*;
- **QS HPV C1**, 1 x 0,2 ml (HPV DNA C+ 16, 18, 45 and human DNA)\*\*;
- **QS HPV C2**, 1 x 0,2 ml (HPV DNA C+ 16, 18, 45 and human DNA)\*\*;

Contains reagents for 108 samples.

\* *must be used in the isolation procedure as Negative Control of Extraction.*

\*\* *Standards' concentration is specific for every lot (reported on the Data Card)*

## **MATERIALS REQUIRED BUT NOT PROVIDED**

### **Zone 1: sample preparation:**

- Biological cabinet
- Desktop microcentrifuge for “eppendorf” type tubes
- 65°C ± 2°C dry heat block
- Vortex mixer
- Pipettes with sterile, RNase-free filters tips
- 1,5 ml polypropylene sterile tubes
- Disposable gloves, powderless
- Tube racks

### **Zone 2: Real Time amplification:**

- Real Time Thermalcycler with 5 fluorescence channels
- PCR Tubes
- Workstation
- Pipettes with sterile, RNase-free filters tips
- Tube racks

## **STORAGE INSTRUCTIONS**

**HPV 14 Screening & 16,18,45 Typing Real-TM Quant** must be stored at -20°C. The kit can be shipped at 2-8°C for 3-4 days but should be stored at -20°C immediately on receipt.

## **STABILITY**

**HPV 14 Screening & 16,18,45 Typing Real-TM Quant** is stable up to the expiration date indicated on the kit label. The product will maintain performance through the control date printed on the label. Exposure to light, heat or humidity may affect the shelf life of some of the kit components and should be avoided. Repeated thawing and freezing of these reagents should be avoided, as this may reduce the sensitivity. Components stored under conditions other than those stated on the labels may not perform properly and may adversely affect the assay results.

## WARNINGS AND PRECAUTIONS



### ***In Vitro* Diagnostic Medical Device**

For *In Vitro* Diagnostic Use Only

The user should always pay attention to the following:

- Use sterile pipette tips with aerosol barriers and use new tip for every procedure.
- Store extracted positive material (samples, controls and amplicons) away from all other reagents and add it to the reaction mix in a separate area.
- Thaw all components thoroughly at room temperature before starting an assay.
- When thawed, mix the components and centrifuge briefly.
- Use disposable gloves, laboratory coats and eye protection when handling specimens and reagents. Thoroughly wash hands afterwards.
- Do not eat, drink, smoke, apply cosmetics, or handle contact lenses in laboratory work areas.
- Do not use a kit after its expiration date.
- Dispose of all specimens and unused reagents in accordance with local authorities' regulations.
- Specimens should be considered potentially infectious and handled in a biological cabinet in accordance with appropriate biosafety practices.
- Clean and disinfect all sample or reagent spills using a disinfectant such as 0.5% sodium hypochlorite, or other suitable disinfectant.
- Avoid sample or reagent contact with the skin, eyes, and mucous membranes. If skin, eyes, or mucous membranes come into contact, rinse immediately with water and seek medical advice immediately.
- Material Safety Data Sheets (MSDS) are available on request.
- Use of this product should be limited to personnel trained in the techniques of DNA amplification.
- The laboratory process must be one-directional, it should begin in the Extraction Area and then move to the Amplification and Detection Areas. Do not return samples, equipment and reagents to the area in which the previous step was performed.



Some components of this kit contain sodium azide as a preservative. Do not use metal tubing for reagent transfer.

## SAMPLE COLLECTION, STORAGE AND TRANSPORT

**HPV 14 Screening & 16,18,45 Typing Real-TM Quant** can analyze DNA extracted with **DNA-Sorb-A** (REF K-1-1/A) from:

- *Cervical swabs:*
  - Remove excess mucus from the cervical os and surrounding ectocervix using a cotton or polyester swab. Discard this swab.
  - Insert the Sampling Cervical Brush 1.0-1.5 centimeters into the cervical os until the largest bristles touch the ectocervix. Do not insert brush completely into the cervical canal. Rotate brush 3 full turns in a counterclockwise direction, remove from the canal.
  - Insert brush into the nuclease-free 2,0 ml tube with 0,3 mL of Transport medium (Sacace). Vigorously agitate brush in medium for 15-20 sec.
  - Snap off shaft at scored line, leaving brush end inside tube.
- *Liquid-based cytology samples* (Cytoscreen, PreservCyt) (recommended DNA-Sorb-D (REF) K-1-8/100 not included in this kit, but can be ordered separately)

It is recommended to process samples immediately after collection. Store samples at 2–8 °C for no longer than 24 hours, or freeze at -20/-80°C. Transportation of clinical specimens must comply with country, federal, state and local regulations for the transport of etiologic agents.

## DNA ISOLATION

Any commercial RNA/DNA isolation kit, if IVD-CE validated for the specimen types indicated herein at the “SAMPLE COLLECTION, STORAGE AND TRANSPORT” paragraph, could be used.

Sacace Biotechnologies recommends to use the following kits:

- ⇒ **DNA-Sorb-A** (Sacace, REF K-1-1/A);
- ⇒ **SaMag STD DNA Extraction kit** (Sacace, REF SM007).

Please carry out DNA extraction according to the manufacture’s instruction.

## REAGENT PREPARATION

### Protocol:

1. Thaw the tube with **PCR-mix-1-FRT HPV 14**. Vortex the tubes with **PCR-mix1-FRT HPV 14, PCR-mix-2 buffer, TaqF DNA Polymerase** and then centrifuge briefly.

Take the required number of tubes/strips for amplification of the DNA obtained from clinical and control samples.

2. For N reactions (including clinical samples, controls and for quantitative analysis also calibrators), add to a new tube:

**10x(N+1) µl of PCR-mix-1-FRT HPV 14,**

**5,0x(N+1) µl of PCR-mix-2 buffer,**

**0,5x(N+1) µl of TaqF DNA Polymerase.**

Vortex the tube, then centrifuge it briefly. Transfer **15 µl** of the prepared mixture to each tube.

3. Add **10 µl** of **DNA samples** extracted from test or control samples to the prepared tubes.
4. Carry out the control amplification reactions (for qualitative analysis add only **QS HPV C2** in 1 PCR reaction tube and **Negative Control** in another PCR reaction tube):

**C1** – Add **10 µl** of **QS HPV C1 HPV 16,18,45 / Glob** to 2 PCR reaction tubes

**C2** – Add **10 µl** of **QS HPV C2 HPV 16,18,45 / Glob** to 2 PCR reaction tubes

**C-** – Add **10 µl** of **the sample extracted from the Negative Control reagent**

1. Program position of the samples and enter the concentrations of the Quantitative Standards (reported in the Quant Data Card) in the Joe (Yellow)/HEX, Fam (Green), Rox(Orange), Cy5 (Red) and Cy5.5 (Crimson) channels in order to generate Standard curves. Use name “Unknown” for the wells that contain samples, C1, C2 for “Standards” and “-” for Negative Controls.

## Amplification

1. Create a temperature profile on your instrument<sup>1</sup> as follows:

Step	Temperature, °C	Time	Fluorescence detection	Cycles
Hold	95	15 min	–	1
Cycling	95	10 s	–	45
	60	25 s	Fluorescence detection*	

<sup>1</sup> For example, SaCycler-96™ (Sacace), Rotor-Gene™ 6000/Q (Corbett Research, Qiagen) CFX96 (BioRad), iQ5 (BioRad), Mx3005P™ (Agilent), ABI® 7500 Real Time PCR (Applied Biosystems)

\* detection on Fam (Green), Joe (Yellow)/Hex, Rox (Orange), Cy5 (Red) and only for 5-channels instruments also Cy5.5 (Crimson/Quasar 705)

## INSTRUMENT SETTINGS

### Rotor-type instruments

Channel	Calibrate/Gain Optimisation...	Threshold	More Settings/ Outlier Removal	Slope Correct
FAM/Green	from 5 FI to 10 FI	0.03	10 %	On
JOE/Yellow	from 5 FI to 10 FI	0.03	10 %	On
Rox (Orange)	from 5 FI to 10 FI	0.03	7 %	On
Cy5 (Red)	from 5 FI to 10 FI	0.03	10 %	On
Cy5.5 (Crimson)	from 5 FI to 10 FI	0.03	10 %	On

### Plate-type instruments

The threshold line should cross only sigmoid curves of signal accumulation of positive samples and should not cross the baseline; otherwise, the threshold level should be raised. Set the threshold at a level where fluorescence curves are linear and do not cross curves of the negative samples.

**Data Analysis:**

The experiment may be considered valid if:

- the Negative Controls haven't any positive fluorescence signal;
- the standards have positive signals in all channels (*Fam, Joe/Hex, Rox, Cy5, Cy5.5\**)

FAM	JOE	ROX	Cy5	Cy5.5 *
DNA of HPV genotype 16	DNA of HPV genotype 18	DNA of HPV (genotypes 16,18,31,33,35,39,45,51,52,56,58,59,66,68)	IC β-globin DNA	DNA of HPV genotype 45

*\*only for 5-channels instruments like SaCycler-96 5X, RotorGene, CFX-96 (4-channels instrument will not be able to genotype HPV 45)*

**Qualitative analysis:**

**Results interpretation of test samples**

FAM	JOE	ROX	Cy5	Cy5.5	Result
<i>Ct value</i>					
Absent or Ct > 36	Absent or Ct > 36	Ct < 36	Ct < 25	Absent or Ct > 36	HPV DNA positive
Ct < 36	Absent or Ct > 36	Ct < 36	Ct < 25	Absent or Ct > 36	HPV DNA positive (including DNA of genotype 16)
Absent or Ct > 36	Ct < 36	Ct < 36	Ct < 25	Absent or Ct > 36	HPV DNA positive (including DNA of genotype 18)
Absent or Ct > 36	Absent or Ct > 36	Ct < 36	Ct < 25	Ct < 36	HPV DNA positive (including DNA of genotype 45)
Absent or Ct > 36	Absent or Ct > 36	Absent or Ct > 36	Ct < 25	Absent or Ct > 36	HPV DNA negative
Absent or Ct > 36	Absent or Ct > 36	Absent or Ct > 36	Absent or Ct > 25	Absent or Ct > 36	Invalid**

**Quantitative analysis:**

Calculate the concentration of HPV DNA using the following formula:

$$\lg \left( \frac{\text{number of HPV DNA copies/reaction}}{\text{number of human DNA copies/reaction}} \times 2 \cdot 10^5 \right) = \lg (\text{HPV DNA copies} / 10^5 \text{ cells})$$

\*\* Invalid samples should be repeated starting from the extraction step



**RESULTS INTERPRETATION:**

Result Ig (HPV per 100,000 human cells)	Interpretation
<3	Clinically insignificant value
3–5	Clinically significant value. Dysplasia cannot be excluded; risk of dysplasia development
>5	Clinically significant, increased value. High probability of dysplasia
Integration (only for genotypes 16, 18 and 45)	Identification of E6 area in the absence of E1/E2 area indirectly suggests the probability of viral integration into the human DNA.

The result is **invalid** if the concentration value of human DNA (obtained for samples in the channel for Cy5 fluorophore) is less than  $10^3$  copies/reaction and the calculated concentration values are absent in the channels for FAM, JOE, ROX, Cy5.5\* fluorophores. It is necessary to repeat the PCR analysis of this sample starting from DNA extraction stage. If human DNA is absent in the test sample, it is recommended to repeat biological material sampling and PCR-analysis.

**The results of the quantitative analysis is considered reliable only if the results obtained for Calibrators and Negative Control of extraction are correct (see Table)**


**Results for controls**

Control	Stage for control	Ct value in the channel for fluorophore				
		FAM	JOE	ROX	Cy5	Cy5.5
C–	DNA extraction	Absent	Absent	Absent	Absent	Absent
C1, C2	PCR	<b>Defined</b>	<b>Defined</b>	<b>Defined</b>	<b>Defined</b>	<b>Defined</b>


\*Cy5.5 is only for 5-channels instruments like SaCycler-96 5X, RotorGene, CFX-96

The results can be calculated automatically using the program in Microsoft® Excel format supplied with the kit.


1. Open the program “HPV 14 Screening & 16,18,45 Typing Quant 4x” and in the window “Security Warning” click on the button “Enable Macros” (Security level of the Microsoft® Excel must be selected as Medium (Tools→Macro→Security→Medium)).
2. Copy with the right button of the mouse the names of the samples from the column “Name” and paste them in the column “Name” of the program “HPV 14 Screening & 16,18,45 Typing Quant 4x”.
3. Copy in the same way the Ct values from the channel FAM (Green) and paste them in the correspond column of the program. Repeat the same procedure for all channels. Standards must be named as C1, C2 and Negative controls must be marked as “C-”.
4. At the top right of the window insert in the table “DNA Calibrator values” the concentrations of the Quantitative standards reported in the Quant Data Card.
5. Click on the buttons “Mark unnamed as empty” and “Calculate”.
6. Save the file with a new name.




*Fam channel – HPV 16 DNA*




*Cy5 channel - IC  $\beta$ -globin DNA*



*Joe channel – HPV 18 DNA*



*Rox channel – HPV DNA (genotypes 16,18,31,33,35,39,45,51,52,56,58,59,66,68)*



*Cy5.5 channel – HPV 45 DNA*

## PERFORMANCE CHARACTERISTICS

### Analytical sensitivity

The kit **HPV 14 Screening & 16,18,45 Typing Real-TM Quant** allows to detect *Human Papillomavirus* DNA in 100% of the tests with a sensitivity of not less than 1000 copies/ml.

Biological material	Transport medium	DNA-extraction kit	HPV genotype	Analytical sensitivity, copies/ml
Swab of vaginal mucosa, scraping of membrane mucosa of cervix uteri and urethra, endocervical scraping	Transport medium for swabs or Transport Medium with Mucolytic Agent	DNA-sorb-A	16	1x10 <sup>3</sup>
			18	1x10 <sup>3</sup>
			31	1x10 <sup>3</sup>
			33	1x10 <sup>3</sup>
			35	1x10 <sup>3</sup>
			39	1x10 <sup>3</sup>
			45	1x10 <sup>3</sup>
			51	1x10 <sup>3</sup>
			52	1x10 <sup>3</sup>
			56	1x10 <sup>3</sup>
			58	1x10 <sup>3</sup>
			59	1x10 <sup>3</sup>
			66	1x10 <sup>3</sup>
68	1x10 <sup>3</sup>			



The claimed sensitivity is achieved only when biomaterial pretreatment is carried out in accordance with the section *SAMPLE COLLECTION, STORAGE AND TRANSPORT*.

The analytical sensitivity for each microorganism is preserved in the presence of high DNA concentrations of other analyte microorganism (10<sup>8</sup> GE/ml).

### Linear measurement range and detection limit

Biological material	Transport medium	DNA-extraction kit	HPV genotype	Detection limit, copies/ml	Linear measurement range, copies/ml
Swab of vaginal mucosa, scraping of membrane mucosa of cervix uteri and urethra, endocervical scraping	Transport medium for swabs or Transport Medium with Mucolytic Agent	DNA-sorb-A	16	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			18	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			31	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			33	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			35	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			39	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			45	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			51	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			52	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			56	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			58	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			59	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
			66	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>
68	1x10 <sup>3</sup>	1x10 <sup>3</sup> – 1x10 <sup>8</sup>			



The claimed values of characteristic are achieved only when biomaterial pretreatment is carried out in accordance with the section *SAMPLE COLLECTION, STORAGE AND TRANSPORT*.

### Analytical specificity

The analytical specificity of the primers and probes was validated with negative samples. They did not generate any signal with the specific *Human Papillomavirus* primers and probes. The specificity of the kit **HPV 14 Screening & 16,18,45 Typing Real-TM Quant** was 100%. The potential cross-reactivity of the kit was tested against the group control. It was not observed any cross-reactivity with other pathogens. Tested pathogens were *Neisseria gonorrhoeae*, *Chlamidia trachomatis*, *Gardnerella vaginalis*, *Mycoplasma genitalium*, *Trichomonas vaginalis*, *Atopobium vaginae*, *Ureaplasma sp.*, *Mycoplasma hominis*, *Ureaplasma parvum*, *Cytomegalovirus*, *Streptococcus agalactiae*, *HSV I*, *HSV II*, *EBV*, *Varicella-Zoster virus*, *Streptococcus pyogenes*, *Candida*, *Human papillomavirus* of low and unknown risk (genotypes 6, 11, 67, 70, 84, 81, 82, 62, 72, 73).

**Target region:** E2, E6, E7

### Reproducibility

HPV genotype	Initial concentration value, lg copies/ml	Number of repeats	Average concentration value, lg copies/ml	Standard deviation (SD)	The coefficient of variation (CV), %
16	3,2 - 4	40	3,55	0,04	1,19
18		40	3,94	0,07	1,71
31		40	3,73	0,09	2,55
33		40	3,68	0,13	3,44
35		40	3,45	0,15	4,30
39		40	3,66	0,12	3,32
45		40	3,65	0,06	1,58
51		40	3,50	0,14	4,09
52		40	3,64	0,11	2,96
56		40	3,78	0,13	3,36
58		40	3,79	0,11	2,86
59		40	3,46	0,09	2,46
66		40	3,88	0,09	2,35
68		40	3,60	0,13	3,53

### Repeatability

HPV genotype	Initial concentration value, lg copies/ml	Number of repeats	Average concentration value, lg copies/ml	Standard deviation (SD)	The coefficient of variation (CV), %
16	3,2 - 4	40	3,56	0,04	1,18
18		40	3,91	0,06	1,50
31		40	3,78	0,05	1,19
33		40	3,57	0,03	0,95
35		40	3,33	0,07	2,18
39		40	3,56	0,08	2,19
45		40	3,61	0,03	0,84
51		40	3,37	0,04	1,09
52		40	3,58	0,10	2,76
56		40	3,68	0,08	2,13
58		40	3,70	0,08	2,07
59		40	3,41	0,08	2,40
66		40	3,81	0,06	1,61
68		40	3,51	0,07	2,12

The accuracy was determined by testing the quality control samples with the concentration of at least  $5 \times 10^4$  copies/ml.

## Accuracy

HPV genotype	Number of repeats	Average concentration value, lg copies/ml	Defined value lg copies/ml	Bias (B), %
16	40	3,55	3,72	4,63
18	40	3,94	3,68	6,62
31	40	3,73	3,67	1,63
33	40	3,68	3,66	0,59
35	40	3,45	3,78	9,49
39	40	3,66	3,77	3,01
45	40	3,65	3,71	1,80
51	40	3,50	3,62	3,60
52	40	3,64	3,72	2,37
56	40	3,78	3,76	0,33
58	40	3,79	3,81	0,48
59	40	3,46	3,69	6,68
66	40	3,88	3,61	5,37
68	40	3,60	3,81	5,60

## Diagnostic characteristics

Diagnostic characteristics of **HPV 14 Screening & 16,18,45 Typing Real-TM Quant** PCR kit were determined according to international requirements for validation of new tests for *HPV* DNA detection.

**Diagnostic sensitivity** of *HPV* test for CIN2+ detection should be not at least 90% of sensitivity of Hybrid Capture 2 method (HC2) (Digene hc2 High-Risk *HPV* DNA Test) according to international requirements for validation of new tests for *HPV* DNA. This means that relative sensitivity is at least 90% and the samples should be histologically confirmed (CIN2 as a minimum). At least 60 samples should be tested using two *HPV* tests.

**Diagnostic specificity** of *HPV* test for CIN2+ detection should be not at least 98% of specificity of Hybrid Capture 2 method (HC2) (Digene hc2 High-Risk *HPV* DNA Test) according to international requirements for validation of new tests for *HPV* DNA. The sampling should include of at least 800 samples obtained from women over age 30 without cytologically/histologically confirmed CIN2.

The 888 samples (scrapings of membrane mucosa of cervix uteri, endocervical scrapings) were studied to determine diagnostic sensitivity and specificity of the kit. The 74 of these samples are with histologically confirmed diagnosis of CIN2+ (26 – CIN2, 37 – CIN3, 4 – AIS/ADC, 7 – SCC) and the average age of female patients is 35 years old (from 20 to 65 years old). And 814 of all samples obtained from screening study are with cytologically/histologically confirmed absence of CIN2. An average age of the women is 39 years old (from 30 to 65 years old). HC2 test (Digene *HPV* test) was used as reference method.

Moreover 184 samples of vaginal mucosal swabs obtained from screening study were studied. **HPV 14 Screening & 16,18,45 Typing Real-TM Quant** kit was used as reference method.

**The results of testing HPV 14 Screening & 16,18,45 Typing Real-TM Quant kit in comparison with reference method**

Samples type	Results of HPV 14 Screening & 16,18,45 Typing Real-TM Quant PCR kit	Results of reference method <sup>1</sup>		
		Positive	Negative	
Scraping of membrane mucosa of cervix uteri, endocervical scrapings with hystologically confirmed moderate or severe dysplasia (CIN2+)	74 samples were investigated	Positive	70	3
		Negative	0	1
Scraping of membrane mucosa of cervix uteri, endocervical scrapings with normal cytology or mild dysplasia	814 samples were investigated	Positive	89	13
		Negative	11	701
Swabs of vaginal mucosa	184 samples were investigated	Positive	46	8
		Negative	0	130

**Diagnostic characteristics of HPV 14 Screening & 16,18,45 Typing Real-TM Quant kit**

Samples type	Diagnostic sensitivity <sup>2</sup> , %	Diagnostic specificity <sup>3</sup> , %
Scraping of membrane mucosa of cervix uteri, endocervical scraping	100	98
Swab of vaginal mucosa	100	94

1) Digene hc2 High-Risk HPV DNA Test kit for scrapings of membrane mucosa of cervix uteri, endocervical scrapings and HPV 14 Screening & 16,18,45 Typing Real-TM Quant kit (for swabs of vaginal mucosa) were used as reference method .

2) Relative sensitivity in comparison with applied reference method (Digene HPV test).

3) Relative specificity in comparison with applied reference method (Digene HPV test).

## KEY TO SYMBOLS USED




List Number



Caution!



Lot Number



Contains sufficient  
for <n> tests



For *in Vitro* Diagnostic  
Use



Version



Store at

**NCA**

Negative Control of  
Amplification



Manufacturer

**C-**

Negative control of  
Extraction



Consult instructions for  
use

**C+**

Positive Control of  
Amplification



Expiration Date

**IC**

Internal Control

- \* SaCycler™ is a registered trademark of Sacace Biotechnologies
- \* CFX™ and iQ5™ are registered trademarks of Bio-Rad Laboratories
- \* Rotor-Gene™ is a registered trademark of Qiagen
- \* MX3005P® is a registered trademark of Agilent Technologies
- \* ABI® is a registered trademark of Applied Biosystems



**Sacace Biotechnologies Srl**  
via Scalabrini, 44 – 22100 – Como – Italy Tel +390314892927 Fax +390314892926  
mail: [info@sacace.com](mailto:info@sacace.com) web: [www.sacace.com](http://www.sacace.com)