



ZeptoBind

Catalog Number: 0801184L **Unit Size:** 1000 mL
Catalog Number: 0801184 **Unit Size:** 60 mL

PRODUCT CHARACTERISTICS

Preparation of microwell plates involve binding the antigen or antibody to the plate, blocking nonspecific binding sites, then coating the plate with a stabilizing reagent to allow for long term storage. ZeptoBind is specifically formulated to aid in the binding of antigens or antibodies to microwell plates. The reagent is supplied in a liquid, ready-to-use format.



RECOMMENDED USAGE

The concentration of antibody or antigen will greatly affect the quality of the assay being developed. We recommend titrating the amount of antigen or antibody to be used for coating the wells. For most antibodies or antigens, the optimal coating concentration is in the range of 1 to 10 $\mu\text{g/mL}$ when using ZeptoBind. If the optimal concentration is unknown and it is not practical to perform a titration to determine this concentration, we recommend using 10 $\mu\text{g/mL}$, which works well for most applications. A complete discussion of optimization procedures can be found in Harlow and Lane, "Antibodies, A Laboratory Manual", Cold Spring Harbor Laboratory, 1988.

Dilute antigen or antibody in ZeptoBind. Add 100 μL of the suspension to each microwell and incubate overnight at 4°C. Aspirate the remaining solution. It is recommended that after binding the antibody or antigen to the plate, that a blocking reagent such as ZeptoBlock (Catalog # 0801185) and a stabilizing reagent such as ZeptoCoat (Catalog # 0801179) are applied. After binding the antigen or antibody to the plate, add 200 μL of ZeptoBlock to each well. Incubate for 2 hours at room temperature. Aspirate remaining solution then add 100 μL of ZeptoCoat to each well. Incubate for 1 hour at room temperature. Aspirate the remaining solution and allow the plate to dry at room temperature for a minimum of 4 hours or overnight. Seal in a bag with a desiccant and store at 2-8 °C.

STORAGE AND STABILITY

Store at 2-8 °C. ZeptoBind contains no preservatives. Use aseptic technique when removing contents to avoid possible contamination. Best if used before expiration date (see product label).

REF	Catalog Number		Temperature Limitation		Expiration Date
LOT	Lot Number	RUO	For Research Use Only		

PI0801184

Rev. No./Replaces: 2 / 03/2015

PCA No.: 17-069 16-191

ZeptoMetrix Corporation • 878 Main Street, Buffalo, NY 14202 • 800-274-5487 • www.zeptometrix.com

FOR RESEARCH USE ONLY. NOT FOR *in vitro* DIAGNOSTIC OR CLINICAL USE

This Package Insert is provided for product evaluation purposes only and is not intended to be used in place of the Package Insert shipped with the product.